

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102618\
 Data File : VX005602.D
 Acq On : 26 Oct 2018 14:02
 Operator : JC/MD
 Sample : J5665-02 10X
 Misc : 5.64g/10ml/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DL-06B

Quant Time: Oct 27 05:15:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	199285	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	293332	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	286731	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	158661	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	124524	51.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.38%	
35) Dibromofluoromethane	5.49	113	103563	52.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.28%	
50) Toluene-d8	8.71	98	378378	53.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.04%	
62) 4-Bromofluorobenzene	11.14	95	144686	53.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.36%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	7.46	83	23189	7.377	ug/l	94
67) Ethyl Benzene	10.25	91	814344	81.064	ug/l	100
68) m/p-Xylenes	10.36	106	1138408	292.090	ug/l	99
69) o-Xylene	10.70	106	88751	24.204	ug/l	99
73) Isopropylbenzene	11.02	105	107120	11.045	ug/l	100
78) n-propylbenzene	11.36	91	313093	28.232	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	459119	55.337	ug/l	100
84) 1,2,4-Trimethylbenzene	11.81	105	1434331	170.577	ug/l	99
85) sec-Butylbenzene	11.94	105	60051	6.085	ug/l	99
86) p-Isopropyltoluene	12.07	119	72336	8.200	ug/l	97
89) n-Butylbenzene	12.38	91	563568	70.764	ug/l #	24
95) Naphthalene	13.83	128	2025786	186.955	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102618\
Data File : VX005602.D
Acq On : 26 Oct 2018 14:02
Operator : JC/MD
Sample : J5665-02 10X
Misc : 5.64g/10ml/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DL-06B

Quant Time: Oct 27 05:15:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

