

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX102618\
 Data File : VX005603.D
 Acq On : 26 Oct 2018 14:29
 Operator : JC/MD
 Sample : J5675-05ME
 Misc : 6.03g/5ml/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-13ME

Quant Time: Oct 29 09:20:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	210585	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	6.86	114	316585	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	274534	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	143325	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	128210	49.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.76%	
35) Dibromofluoromethane	5.51	113	101144	47.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.36%	
50) Toluene-d8	8.71	98	360182	46.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.52%	
62) 4-Bromofluorobenzene	11.14	95	124097	42.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.32%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102618\
Data File : VX005603.D
Acq On : 26 Oct 2018 14:29
Operator : JC/MD
Sample : J5675-05ME
Misc : 6.03g/5ml/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-13ME

Quant Time: Oct 29 09:20:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

