

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102618\
 Data File : VX005613.D
 Acq On : 26 Oct 2018 18:56
 Operator : JC/MD
 Sample : J5696-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2018.10.25-SED

Quant Time: Oct 29 09:28:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	262158	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	376896	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	321218	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	153108	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	153338	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
35) Dibromofluoromethane	5.51	113	117508	46.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.10%	
50) Toluene-d8	8.71	98	430332	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
62) 4-Bromofluorobenzene	11.14	95	144427	41.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.42%	
Target Compounds						
16) Acetone	2.45	43	9820	5.807	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102618\
Data File : VX005613.D
Acq On : 26 Oct 2018 18:56
Operator : JC/MD
Sample : J5696-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2018.10.25-SED

Quant Time: Oct 29 09:28:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

