

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102618\
 Data File : VX005619.D
 Acq On : 26 Oct 2018 21:35
 Operator : JC/MD
 Sample : J5676-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-1018-S-TCLP-GRID40

Quant Time: Oct 27 05:52:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	252976	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	351975	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	312608	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	151976	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	144756	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
35) Dibromofluoromethane	5.51	113	124123	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
50) Toluene-d8	8.71	98	408993	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
62) 4-Bromofluorobenzene	11.14	95	141327	43.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.40%	
Target Compounds						
16) Acetone	2.45	43	14972	9.175	ug/l	97
20) Methylene Chloride	2.85	84	2154977	788.522	ug/l	89
25) 2-Butanone	4.70	43	3542	1.425	ug/l	96
43) Isopropyl Acetate	6.46	43	108893	16.671	ug/l	97
68) m/p-Xylenes	10.36	106	5278	1.242	ug/l	95
95) Naphthalene	13.83	128	27399	2.640	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102618\
Data File : VX005619.D
Acq On : 26 Oct 2018 21:35
Operator : JC/MD
Sample : J5676-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-1018-S-TCLP-GRID40

Quant Time: Oct 27 05:52:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

