

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102618\
 Data File : VX005632.D
 Acq On : 27 Oct 2018 04:14
 Operator : JC/MD
 Sample : J5696-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DL-SW-02

Quant Time: Oct 29 12:24:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251576	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	371573	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	325399	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	160780	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	152870	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
35) Dibromofluoromethane	5.50	113	117879	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	
50) Toluene-d8	8.71	98	432280	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
62) 4-Bromofluorobenzene	11.14	95	149606	43.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.64%	
Target Compounds						
16) Acetone	2.45	43	7790	4.801	ug/l	99
30) Chloroform	5.21	83	9451	2.019	ug/l	95
44) Trichloroethene	7.21	130	1030	0.366	ug/l	92
47) Bromodichloromethane	7.90	83	1753	0.503	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102618\
Data File : VX005632.D
Acq On : 27 Oct 2018 04:14
Operator : JC/MD
Sample : J5696-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DL-SW-02

Quant Time: Oct 29 12:24:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

