

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102618\
 Data File : VX005640.D
 Acq On : 27 Oct 2018 07:47
 Operator : JC/MD
 Sample : J5692-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FT-PZ459I-20181025

Quant Time: Oct 29 12:37:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	230702	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	318609	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	290220	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186729	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	140800	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
35) Dibromofluoromethane	5.50	113	117963	54.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.36%	
50) Toluene-d8	8.71	98	384094	49.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.10%	
62) 4-Bromofluorobenzene	11.14	95	129932	44.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.76%	
Target Compounds						
24) 1,1-Dichloroethane	3.70	63	3004	0.655	ug/l #	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102618\
Data File : VX005640.D
Acq On : 27 Oct 2018 07:47
Operator : JC/MD
Sample : J5692-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FT-PZ459I-20181025

Quant Time: Oct 29 12:37:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

