

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102618\
 Data File : VX005644.D
 Acq On : 27 Oct 2018 09:34
 Operator : JC/MD
 Sample : J5204-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-01-102518-A

Quant Time: Oct 29 02:17:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	229048	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	362570	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	302367	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	177469	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	143372	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
35) Dibromofluoromethane	5.49	113	119924	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
50) Toluene-d8	8.71	98	397152	45.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.04%	
62) 4-Bromofluorobenzene	11.13	95	156823	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
Target Compounds						
8) Diethyl Ether	2.19	74	1598	1.018	ug/l	92
16) Acetone	2.45	43	31990	21.652	ug/l	99
20) Methylene Chloride	2.86	84	3671	1.484	ug/l	93
25) 2-Butanone	4.71	43	2770	1.231	ug/l #	78
43) Isopropyl Acetate	6.46	43	160270	23.820	ug/l	99
95) Naphthalene	13.83	128	63600	5.247	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102618\
Data File : VX005644.D
Acq On : 27 Oct 2018 09:34
Operator : JC/MD
Sample : J5204-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
SU-01-102518-A

Quant Time: Oct 29 02:17:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

