

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102621\
 Data File : VX024941.D
 Acq On : 26 Oct 2021 13:13
 Operator : JC/MD
 Sample : VIBLK601
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK601

Quant Time: Oct 27 03:02:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 27 02:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	67369	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	62699	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	29267	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	7063	3.571	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.400%
7) Chloroethane-d5	1.673	69	7717	4.331	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.600%
11) 1,1-Dichloroethene-d2	2.307	63	21609	2.931	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.600%#
20) 2-Butanone-d5	4.483	46	38359	53.047	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.100%
24) Chloroform-d	5.062	84	47594	4.999	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.000%
26) 1,2-Dichloroethane-d4	5.964	65	25565	4.913	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
32) Benzene-d6	5.983	84	72286	4.948	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	7.318	67	22519	4.832	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.600%
41) Toluene-d8	8.653	98	69083	4.716	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.400%
43) trans-1,3-Dichloroprop...	8.958	79	9567	4.472	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.400%
46) 2-Hexanone-d5	9.391	63	40090	52.070	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.140%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	18655	5.237	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.800%
66) 1,2-Dichlorobenzene-d4	12.323	152	25788	5.229	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.600%
Target Compounds						Qvalue
16) Methylene chloride	2.794	84	1082	0.225	ug/L	# 68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102621\
Data File : VX024941.D
Acq On : 26 Oct 2021 13:13
Operator : JC/MD
Sample : VIBLK601
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK601

Quant Time: Oct 27 03:02:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 27 02:59:04 2021
Response via : Initial Calibration

