

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX102621\  
 Data File : VX024947.D  
 Acq On : 26 Oct 2021 15:36  
 Operator : JC/MD  
 Sample : M4304-10  
 Misc : 25.0mL/MSVOA\_X/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 P069-TW001-01

Manual Integrations  
 APPROVED

MMDadoda  
 10/27/2021 1:36:14 PM

Quant Time: Oct 27 03:03:57 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\SFAMXTR102221WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Wed Oct 27 02:59:04 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|--------|----------|
| -----                         |        |       |          |          |        |          |
| Internal Standards            |        |       |          |          |        |          |
| 1) 1,4-Difluorobenzene        | 6.763  | 114   | 69053    | 5.000    | ug/L   | 0.00     |
| 28) Chlorobenzene-d5          | 10.055 | 117   | 62145    | 5.000    | ug/L   | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 12.030 | 152   | 28558    | 5.000    | ug/L   | 0.00     |
|                               |        |       |          |          |        |          |
| System Monitoring Compounds   |        |       |          |          |        |          |
| 4) Vinyl Chloride-d3          | 1.368  | 65    | 6556     | 3.234    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =      | 64.600%  |
| 7) Chloroethane-d5            | 1.672  | 69    | 7245     | 3.966    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =      | 79.400%  |
| 11) 1,1-Dichloroethene-d2     | 2.313  | 63    | 20667    | 2.735    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =      | 54.600%# |
| 20) 2-Butanone-d5             | 4.471  | 46    | 33768    | 45.560   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =      | 91.120%  |
| 24) Chloroform-d              | 5.068  | 84    | 44656    | 4.576    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =      | 91.600%  |
| 26) 1,2-Dichloroethane-d4     | 5.958  | 65    | 23101    | 4.331    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =      | 86.600%  |
| 32) Benzene-d6                | 5.983  | 84    | 65562    | 4.528    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =      | 90.600%  |
| 36) 1,2-Dichloropropane-d6    | 7.312  | 67    | 20399    | 4.416    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =      | 88.400%  |
| 41) Toluene-d8                | 8.653  | 98    | 62818    | 4.327    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =      | 86.600%  |
| 43) trans-1,3-Dichloroprop... | 8.952  | 79    | 8945     | 4.219    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =      | 84.400%  |
| 46) 2-Hexanone-d5             | 9.390  | 63    | 36128    | 47.342   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =      | 94.680%  |
| 56) 1,1,2,2-Tetrachloroeth... | 11.195 | 84    | 17265    | 4.890    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =      | 97.800%  |
| 66) 1,2-Dichlorobenzene-d4    | 12.323 | 152   | 23900    | 4.966    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =      | 99.400%  |
|                               |        |       |          |          |        |          |
| Target Compounds              |        |       |          |          | Qvalue |          |
| 13) Acetone                   | 2.398  | 43    | 1102m    | 1.982    | ug/L   |          |
| -----                         |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX102621\  
Data File : VX024947.D  
Acq On : 26 Oct 2021 15:36  
Operator : JC/MD  
Sample : M4304-10  
Misc : 25.0mL/MSVOA\_X/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
P069-TW001-01

Manual Integrations  
APPROVED

MMDadoda  
10/27/2021 1:36:14 PM

Quant Time: Oct 27 03:03:57 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\SFAMXTR102221WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Wed Oct 27 02:59:04 2021  
Response via : Initial Calibration

