

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102621\
 Data File : VX024946.D
 Acq On : 26 Oct 2021 15:12
 Operator : JC/MD
 Sample : M4304-09
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P068-TW001-01

Quant Time: Oct 27 03:03:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 27 02:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	75152	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	68755	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	31429	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	6841	3.101	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.000%
7) Chloroethane-d5	1.673	69	7624	3.835	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.800%
11) 1,1-Dichloroethene-d2	2.313	63	19910	2.421	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	48.400%#
20) 2-Butanone-d5	4.477	46	35131	43.552	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.100%
24) Chloroform-d	5.062	84	44609	4.200	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.000%
26) 1,2-Dichloroethane-d4	5.964	65	23582	4.062	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.200%
32) Benzene-d6	5.983	84	66828	4.172	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.400%
36) 1,2-Dichloropropane-d6	7.312	67	21009	4.111	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.200%
41) Toluene-d8	8.653	98	63776	3.971	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.400%
43) trans-1,3-Dichloroprop...	8.958	79	9037	3.852	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.000%
46) 2-Hexanone-d5	9.391	63	36071	42.723	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.440%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	17572	4.498	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	12.323	152	24155	4.561	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102621\
Data File : VX024946.D
Acq On : 26 Oct 2021 15:12
Operator : JC/MD
Sample : M4304-09
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P068-TW001-01

Quant Time: Oct 27 03:03:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 27 02:59:04 2021
Response via : Initial Calibration

