

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
 Data File : VX032405.D
 Acq On : 26 Oct 2022 15:22
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 27 08:31:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	134651	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	213476	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187893	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89502	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87294	46.658	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.320%
35) Dibromofluoromethane	5.391	113	68820	45.195	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.380%
50) Toluene-d8	8.652	98	254088	47.987	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.980%
62) 4-Bromofluorobenzene	11.085	95	89350	43.199	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
Data File : VX032405.D
Acq On : 26 Oct 2022 15:22
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 27 08:31:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

