

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
 Data File : VX032412.D
 Acq On : 26 Oct 2022 18:14
 Operator : JC/MD
 Sample : N5237-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LSP-RB

Quant Time: Oct 27 08:33:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	207954	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180456	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86006	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83396	45.909	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.820%
35) Dibromofluoromethane	5.385	113	67054	45.204	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.400%
50) Toluene-d8	8.653	98	244779	47.457	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.920%
62) 4-Bromofluorobenzene	11.079	95	84944	42.160	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	84.320%
Target Compounds						
16) Acetone	2.380	43	3431	5.197	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
Data File : VX032412.D
Acq On : 26 Oct 2022 18:14
Operator : JC/MD
Sample : N5237-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LSP-RB

Quant Time: Oct 27 08:33:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

