

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
 Data File : VX032414.D
 Acq On : 26 Oct 2022 19:01
 Operator : JC/MD
 Sample : N5260-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2022-10-21

Quant Time: Oct 27 08:33:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120266	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	194260	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168637	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77068	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79472	47.558	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.120%
35) Dibromofluoromethane	5.385	113	62167	44.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.720%
50) Toluene-d8	8.653	98	230396	47.817	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.640%
62) 4-Bromofluorobenzene	11.079	95	78050	41.469	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.940%#
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	622	0.447	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
Data File : VX032414.D
Acq On : 26 Oct 2022 19:01
Operator : JC/MD
Sample : N5260-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2022-10-21

Quant Time: Oct 27 08:33:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

