

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
 Data File : VX032415.D
 Acq On : 26 Oct 2022 19:25
 Operator : JC/MD
 Sample : N5248-08RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-20-20221018RE

Quant Time: Oct 27 08:33:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	118590	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	192683	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78411	47.586	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.180%
35) Dibromofluoromethane	5.391	113	61632	44.842	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.680%
50) Toluene-d8	8.653	98	225280	47.138	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.280%
62) 4-Bromofluorobenzene	11.079	95	77455	41.489	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.980%#
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	54387	35.671	ug/l	99
52) Toluene	8.720	92	2325	0.658	ug/l	96
64) Tetrachloroethene	9.275	164	2278	1.887	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
Data File : VX032415.D
Acq On : 26 Oct 2022 19:25
Operator : JC/MD
Sample : N5248-08RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-20-20221018RE

Quant Time: Oct 27 08:33:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

