

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX102622\
 Data File : VX032406.D
 Acq On : 26 Oct 2022 15:45
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX102622

Quant Time: Oct 27 08:32:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	50.000	57.127	-14.3	101	0.00
3 P	Chloromethane	50.000	54.969	-9.9	101	0.00
4 C	Vinyl Chloride	50.000	54.389	-8.8#	98	0.00
5 T	Bromomethane	50.000	55.539	-11.1	102	0.00
6 T	Chloroethane	50.000	51.291	-2.6	102	0.00
7 T	Trichlorofluoromethane	50.000	52.691	-5.4	99	0.00
8 T	Diethyl Ether	50.000	51.911	-3.8	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	52.943	-5.9	102	0.00
10 T	Methyl Iodide	50.000	49.893	0.2	94	0.00
11 T	Tert butyl alcohol	250.000	258.785	-3.5	99	0.00
12 CM	1,1-Dichloroethene	50.000	51.736	-3.5#	99	0.00
13 T	Acrolein	250.000	253.817	-1.5	100	0.00
14 T	Allyl chloride	50.000	52.814	-5.6	99	0.00
15 T	Acrylonitrile	250.000	255.140	-2.1	101	0.00
16 T	Acetone	250.000	254.953	-2.0	105	0.00
17 T	Carbon Disulfide	50.000	54.271	-8.5	100	0.00
18 T	Methyl Acetate	50.000	51.665	-3.3	99	0.00
19 T	Methyl tert-butyl Ether	50.000	51.505	-3.0	101	0.00
20 T	Methylene Chloride	50.000	49.676	0.6	101	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.135	-2.3	99	0.00
22 T	Diisopropyl ether	50.000	51.171	-2.3	100	0.00
23 T	Vinyl Acetate	250.000	262.521	-5.0	100	0.00
24 P	1,1-Dichloroethane	50.000	51.126	-2.3	99	0.00
25 T	2-Butanone	250.000	256.885	-2.8	101	0.00
26 T	2,2-Dichloropropane	50.000	54.456	-8.9	103	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.811	-3.6	98	0.00
28 T	Bromochloromethane	50.000	47.966	4.1	97	0.00
29 T	Tetrahydrofuran	250.000	257.175	-2.9	100	0.00
30 C	Chloroform	50.000	51.571	-3.1#	100	0.00
31 T	Cyclohexane	50.000	52.984	-6.0	99	0.00
32 T	1,1,1-Trichloroethane	50.000	52.511	-5.0	99	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.834	0.3	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	50.333	-0.7	97	0.00
36 T	1,1-Dichloropropene	50.000	53.264	-6.5	100	0.00
37 T	Ethyl Acetate	50.000	50.128	-0.3	101	0.00
38 T	Carbon Tetrachloride	50.000	53.660	-7.3	99	0.00
39 T	Methylcyclohexane	50.000	55.691	-11.4	101	0.00
40 TM	Benzene	50.000	52.553	-5.1	98	0.00
41 T	Methacrylonitrile	50.000	55.335	-10.7	104	0.00
42 TM	1,2-Dichloroethane	50.000	53.750	-7.5	101	0.00
43 T	Isopropyl Acetate	50.000	54.016	-8.0	100	0.00
44 TM	Trichloroethene	50.000	52.586	-5.2	100	0.00
45 C	1,2-Dichloropropane	50.000	52.340	-4.7#	99	0.00
46 T	Dibromomethane	50.000	52.571	-5.1	100	0.00
47 T	Bromodichloromethane	50.000	52.988	-6.0	99	0.00
48 T	Methyl methacrylate	50.000	52.676	-5.4	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
 Data File : VX032406.D
 Acq On : 26 Oct 2022 15:45
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVX102622

Quant Time: Oct 27 08:32:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1053.212	-5.3	97	0.00
50 S	Toluene-d8	50.000	51.378	-2.8	98	0.00
51 T	4-Methyl-2-Pentanone	250.000	266.177	-6.5	99	0.00
52 CM	Toluene	50.000	53.553	-7.1#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	56.573	-13.1	101	0.00
54 T	cis-1,3-Dichloropropene	50.000	55.517	-11.0	102	0.00
55 T	1,1,2-Trichloroethane	50.000	51.587	-3.2	100	0.00
56 T	Ethyl methacrylate	50.000	55.186	-10.4	99	0.00
57 T	1,3-Dichloropropane	50.000	52.600	-5.2	100	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	271.904	-8.8	98	0.00
59 T	2-Hexanone	250.000	274.979	-10.0	101	0.00
60 T	Dibromochloromethane	50.000	53.236	-6.5	99	0.00
61 T	1,2-Dibromoethane	50.000	53.430	-6.9	100	0.00
62 S	4-Bromofluorobenzene	50.000	53.401	-6.8	99	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	53.392	-6.8	101	0.00
65 PM	Chlorobenzene	50.000	52.833	-5.7	102	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.041	-6.1	100	0.00
67 C	Ethyl Benzene	50.000	53.497	-7.0#	100	0.00
68 T	m/p-Xylenes	100.000	108.729	-8.7	101	0.00
69 T	o-Xylene	50.000	54.389	-8.8	101	0.00
70 T	Styrene	50.000	54.995	-10.0	101	0.00
71 P	Bromoform	50.000	54.336	-8.7	98	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	96	0.00
73 T	Isopropylbenzene	50.000	50.232	-0.5	101	0.00
74 T	N-ethyl acetate	50.000	52.545	-5.1	101	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.324	1.4	101	0.00
76 T	1,2,3-Trichloropropane	50.000	51.196	-2.4	101	0.00
77 T	Bromobenzene	50.000	49.615	0.8	102	0.00
78 T	n-propylbenzene	50.000	50.934	-1.9	101	0.00
79 T	2-Chlorotoluene	50.000	50.150	-0.3	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.816	-3.6	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.038	-4.1	100	0.00
82 T	4-Chlorotoluene	50.000	51.140	-2.3	102	0.00
83 T	tert-Butylbenzene	50.000	51.487	-3.0	102	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.747	-3.5	101	0.00
85 T	sec-Butylbenzene	50.000	52.797	-5.6	102	0.00
86 T	p-Isopropyltoluene	50.000	53.876	-7.8	103	0.00
87 T	1,3-Dichlorobenzene	50.000	51.114	-2.2	103	0.00
88 T	1,4-Dichlorobenzene	50.000	49.958	0.1	102	0.00
89 T	n-Butylbenzene	50.000	54.830	-9.7	104	0.00
90 T	Hexachloroethane	50.000	53.220	-6.4	103	0.00
91 T	1,2-Dichlorobenzene	50.000	51.284	-2.6	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.068	-4.1	100	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.655	-3.3	106	0.00
94 T	Hexachlorobutadiene	50.000	51.081	-2.2	105	0.00
95 T	Naphthalene	50.000	50.990	-2.0	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
Data File : VX032406.D
Acq On : 26 Oct 2022 15:45
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX102622

Quant Time: Oct 27 08:32:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.301	1.4	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6