

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX102622\
 Data File : VX032418.D
 Acq On : 26 Oct 2022 20:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 09:09:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	50.000	54.104	-8.2	91	0.00
3 P	Chloromethane	50.000	54.556	-9.1	96	0.00
4 C	Vinyl Chloride	50.000	54.502	-9.0#	94	0.00
5 T	Bromomethane	50.000	55.201	-10.4	96	0.00
6 T	Chloroethane	50.000	50.607	-1.2	96	0.00
7 T	Trichlorofluoromethane	50.000	51.658	-3.3	92	0.00
8 T	Diethyl Ether	50.000	52.213	-4.4	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.369	-2.7	95	0.00
10 T	Methyl Iodide	50.000	51.080	-2.2	92	0.00
11 T	Tert butyl alcohol	250.000	261.320	-4.5	95	0.00
12 CM	1,1-Dichloroethene	50.000	50.951	-1.9#	92	0.00
13 T	Acrolein	250.000	268.773	-7.5	101	0.00
14 T	Allyl chloride	50.000	51.539	-3.1	92	0.00
15 T	Acrylonitrile	250.000	259.783	-3.9	98	0.00
16 T	Acetone	250.000	252.490	-1.0	99	0.00
17 T	Carbon Disulfide	50.000	53.449	-6.9	94	0.00
18 T	Methyl Acetate	50.000	52.856	-5.7	96	0.00
19 T	Methyl tert-butyl Ether	50.000	50.565	-1.1	94	0.00
20 T	Methylene Chloride	50.000	49.480	1.0	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.618	-1.2	94	0.00
22 T	Diisopropyl ether	50.000	50.087	-0.2	93	0.00
23 T	Vinyl Acetate	250.000	258.316	-3.3	93	0.00
24 P	1,1-Dichloroethane	50.000	50.868	-1.7	93	0.00
25 T	2-Butanone	250.000	255.434	-2.2	96	0.00
26 T	2,2-Dichloropropane	50.000	46.195	7.6	83	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.912	-3.8	94	0.00
28 T	Bromochloromethane	50.000	49.802	0.4	96	0.00
29 T	Tetrahydrofuran	250.000	262.322	-4.9	97	0.00
30 C	Chloroform	50.000	50.742	-1.5#	94	0.00
31 T	Cyclohexane	50.000	51.124	-2.2	91	0.00
32 T	1,1,1-Trichloroethane	50.000	51.727	-3.5	93	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.489	-1.0	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	48.684	2.6	94	0.00
36 T	1,1-Dichloropropene	50.000	49.798	0.4	94	0.00
37 T	Ethyl Acetate	50.000	48.905	2.2	99	0.00
38 T	Carbon Tetrachloride	50.000	50.112	-0.2	93	0.00
39 T	Methylcyclohexane	50.000	50.539	-1.1	92	0.00
40 TM	Benzene	50.000	49.403	1.2	93	0.00
41 T	Methacrylonitrile	50.000	52.979	-6.0	100	0.00
42 TM	1,2-Dichloroethane	50.000	50.301	-0.6	95	0.00
43 T	Isopropyl Acetate	50.000	50.251	-0.5	93	0.00
44 TM	Trichloroethene	50.000	48.036	3.9	92	0.00
45 C	1,2-Dichloropropane	50.000	49.337	1.3#	94	0.00
46 T	Dibromomethane	50.000	48.707	2.6	93	0.00
47 T	Bromodichloromethane	50.000	50.422	-0.8	95	0.00
48 T	Methyl methacrylate	50.000	49.935	0.1	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
 Data File : VX032418.D
 Acq On : 26 Oct 2022 20:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 09:09:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1102.090	-10.2	102	0.00
50 S	Toluene-d8	50.000	49.615	0.8	95	0.00
51 T	4-Methyl-2-Pentanone	250.000	255.312	-2.1	96	0.00
52 CM	Toluene	50.000	49.851	0.3#	93	0.00
53 T	t-1,3-Dichloropropene	50.000	50.872	-1.7	92	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.173	-0.3	92	0.00
55 T	1,1,2-Trichloroethane	50.000	48.794	2.4	95	0.00
56 T	Ethyl methacrylate	50.000	50.986	-2.0	92	0.00
57 T	1,3-Dichloropropane	50.000	49.692	0.6	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	257.505	-3.0	94	0.00
59 T	2-Hexanone	250.000	260.817	-4.3	96	0.00
60 T	Dibromochloromethane	50.000	50.069	-0.1	94	0.00
61 T	1,2-Dibromoethane	50.000	49.296	1.4	93	0.00
62 S	4-Bromofluorobenzene	50.000	49.800	0.4	93	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	47.847	4.3	91	0.00
65 PM	Chlorobenzene	50.000	48.591	2.8	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.435	1.1	94	0.00
67 C	Ethyl Benzene	50.000	49.526	0.9#	93	0.00
68 T	m/p-Xylenes	100.000	99.176	0.8	93	0.00
69 T	o-Xylene	50.000	49.837	0.3	93	0.00
70 T	Styrene	50.000	50.161	-0.3	93	0.00
71 P	Bromoform	50.000	49.855	0.3	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	91	0.00
73 T	Isopropylbenzene	50.000	48.783	2.4	93	0.00
74 T	N-ethyl acetate	50.000	50.773	-1.5	93	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.322	3.4	93	0.00
76 T	1,2,3-Trichloropropane	50.000	49.388	1.2	92	0.00
77 T	Bromobenzene	50.000	48.510	3.0	94	0.00
78 T	n-propylbenzene	50.000	48.450	3.1	91	0.00
79 T	2-Chlorotoluene	50.000	48.384	3.2	92	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.555	0.9	92	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.873	4.3	87	0.00
82 T	4-Chlorotoluene	50.000	48.994	2.0	92	0.00
83 T	tert-Butylbenzene	50.000	48.876	2.2	92	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.544	0.9	92	0.00
85 T	sec-Butylbenzene	50.000	50.173	-0.3	92	0.00
86 T	p-Isopropyltoluene	50.000	50.138	-0.3	91	0.00
87 T	1,3-Dichlorobenzene	50.000	48.461	3.1	92	0.00
88 T	1,4-Dichlorobenzene	50.000	47.685	4.6	92	0.00
89 T	n-Butylbenzene	50.000	50.320	-0.6	91	0.00
90 T	Hexachloroethane	50.000	50.486	-1.0	93	0.00
91 T	1,2-Dichlorobenzene	50.000	48.974	2.1	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	50.188	-0.4	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.454	5.1	92	0.00
94 T	Hexachlorobutadiene	50.000	45.809	8.4	90	0.00
95 T	Naphthalene	50.000	48.858	2.3	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102622\
Data File : VX032418.D
Acq On : 26 Oct 2022 20:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 27 09:09:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	46.506	7.0	90	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6