

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102623\
 Data File : VX038525.D
 Acq On : 26 Oct 2023 18:05
 Operator : JC/MD
 Sample : 05097-01 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22

Quant Time: Oct 27 05:20:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	183634	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	314283	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	322892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	177961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	129294	43.775	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.540%
35) Dibromofluoromethane	5.391	113	114476	50.550	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.100%
50) Toluene-d8	8.647	98	419987	48.533	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.060%
62) 4-Bromofluorobenzene	11.079	95	166649	50.797	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.600%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	14617	7.297	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	30738	12.407	ug/l	80
44) Trichloroethene	7.135	130	1641	0.718	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102623\
Data File : VX038525.D
Acq On : 26 Oct 2023 18:05
Operator : JC/MD
Sample : 05097-01 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Oct 27 05:20:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

