

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102623\
 Data File : VX038530.D
 Acq On : 26 Oct 2023 20:02
 Operator : JC/MD
 Sample : 05097-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01

Quant Time: Oct 27 05:22:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	188026	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	318927	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	325967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	179741	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	128261	42.410	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.820%
35) Dibromofluoromethane	5.385	113	115225	50.140	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.280%
50) Toluene-d8	8.653	98	418784	47.689	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.380%
62) 4-Bromofluorobenzene	11.079	95	166500	50.013	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.020%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4322	3.424	ug/l #	81
20) Methylene Chloride	2.794	84	3458	1.457	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102623\
Data File : VX038530.D
Acq On : 26 Oct 2023 20:02
Operator : JC/MD
Sample : 05097-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01

Quant Time: Oct 27 05:22:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

