

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102623\
 Data File : VX038531.D
 Acq On : 26 Oct 2023 20:25
 Operator : JC/MD
 Sample : 05063-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-5SR

Quant Time: Oct 27 05:23:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	195963	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	333320	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	347223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	190725	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135218	42.900	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.800%
35) Dibromofluoromethane	5.391	113	122248	50.899	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.800%
50) Toluene-d8	8.647	98	446781	48.681	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.360%
62) 4-Bromofluorobenzene	11.079	95	179039	51.457	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.920%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2988	2.271	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102623\
Data File : VX038531.D
Acq On : 26 Oct 2023 20:25
Operator : JC/MD
Sample : 05063-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-5SR

Quant Time: Oct 27 05:23:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

