

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102623\
 Data File : VX038529.D
 Acq On : 26 Oct 2023 19:38
 Operator : JC/MD
 Sample : 05097-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB03

Quant Time: Oct 27 05:22:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	184947	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	313764	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	329156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	184816	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	124908	41.989	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.980%
35) Dibromofluoromethane	5.385	113	113567	50.232	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.460%
50) Toluene-d8	8.653	98	414486	47.977	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.960%
62) 4-Bromofluorobenzene	11.079	95	167896	51.262	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.520%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102623\
Data File : VX038529.D
Acq On : 26 Oct 2023 19:38
Operator : JC/MD
Sample : 05097-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB03

Quant Time: Oct 27 05:22:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 711
Delta R.T. -0.000 min
Lab File: VX038529.D
Acq: 26 Oct 2023 19:38




Instrument :
MSVOA_X
ClientSampleId :
TB03


Tgt Ion:168 Resp: 184947
Ion Ratio Lower Upper
168 100
99 52.1 56.6 85.0#



#33
1,2-Dichloroethane-d4
Concen: 41.989 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX038529.D
Acq: 26 Oct 2023 19:38

Tgt Ion: 65 Resp: 124908
Ion Ratio Lower Upper
65 100
67 52.8 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. 0.000 min
Lab File: VX038529.D
Acq: 26 Oct 2023 19:38




Instrument :
MSVOA_X
ClientSampleId :
TB03


Tgt Ion:114 Resp: 313764
Ion Ratio Lower Upper
114 100
63 18.7 0.0 45.0
88 15.0 0.0 36.6



#35
Dibromofluoromethane
Concen: 50.232 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038529.D
Acq: 26 Oct 2023 19:38

Tgt Ion:113 Resp: 113567
Ion Ratio Lower Upper
113 100
111 102.4 82.9 124.3
192 19.7 13.3 19.9









#50
Toluene-d8
Concen: 47.977 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX038529.D
Acq: 26 Oct 2023 19:38




Instrument :
MSVOA_X
ClientSampleId :
TB03


Tgt Ion: 98 Resp: 414486
Ion Ratio Lower Upper
98 100
100 64.3 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 51.262 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX038529.D
Acq: 26 Oct 2023 19:38

Tgt Ion: 95 Resp: 167896
Ion Ratio Lower Upper
95 100
174 87.8 0.0 135.6
176 83.1 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX038529.D

Acq: 26 Oct 2023 19:38

Instrument :

MSVOA_X

ClientSampleId :

TB03





Tgt Ion:117 Resp: 329156

Ion Ratio Lower Upper

117 100

82 53.2 46.2 69.4

119 33.0 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX038529.D

Acq: 26 Oct 2023 19:38

Tgt Ion:152 Resp: 184816

Ion Ratio Lower Upper

152 100

115 57.9 43.3 129.9

150 157.2 0.0 345.0

