

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102720\
 Data File : VX019153.D
 Acq On : 27 Oct 2020 21:40
 Operator : JC/MD
 Sample : L4516-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FRAC-TANK-2

Quant Time: Oct 28 03:06:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	313724	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	550361	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	526667	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	269609	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	208235	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
35) Dibromofluoromethane	5.46	113	168654	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
50) Toluene-d8	8.70	98	675678	51.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.70%	
62) 4-Bromofluorobenzene	11.12	95	262963	53.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.56%	
Target Compounds						
16) Acetone	2.42	43	4762	3.412	ug/l	100
20) Methylene Chloride	2.84	84	1256	0.358	ug/l	85
30) Chloroform	5.17	83	2298	0.377	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102720\
Data File : VX019153.D
Acq On : 27 Oct 2020 21:40
Operator : JC/MD
Sample : L4516-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-2

Quant Time: Oct 28 03:06:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

