

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102721\
 Data File : VX024964.D
 Acq On : 27 Oct 2021 10:36
 Operator : JC/MD
 Sample : VX1027WBL01
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK609

Quant Time: Oct 28 03:10:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 28 03:09:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	75805	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	66999	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	29783	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	11284	5.070	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.400%
7) Chloroethane-d5	1.672	69	9092	4.534	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.600%
11) 1,1-Dichloroethene-d2	2.306	63	28905	3.484	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.600%
20) 2-Butanone-d5	4.483	46	35136	43.183	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.360%
24) Chloroform-d	5.062	84	50769	4.739	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.800%
26) 1,2-Dichloroethane-d4	5.958	65	27352	4.671	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.400%
32) Benzene-d6	5.977	84	83092	5.323	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.400%
36) 1,2-Dichloropropane-d6	7.312	67	23724	4.763	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.200%
41) Toluene-d8	8.653	98	84495	5.398	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.000%
43) trans-1,3-Dichloroprop...	8.952	79	11889	5.201	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	104.000%
46) 2-Hexanone-d5	9.391	63	38624	46.946	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.900%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	18202	4.782	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.600%
66) 1,2-Dichlorobenzene-d4	12.323	152	29271	5.832	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.600%
Target Compounds						
					Qvalue	
16) Methylene chloride	2.794	84	1430	0.265	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102721\
Data File : VX024964.D
Acq On : 27 Oct 2021 10:36
Operator : JC/MD
Sample : VX1027WBL01
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK609

Quant Time: Oct 28 03:10:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 28 03:09:30 2021
Response via : Initial Calibration

