

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102721\
 Data File : VX024968.D
 Acq On : 27 Oct 2021 12:24
 Operator : JC/MD
 Sample : M4304-14
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK001

Quant Time: Oct 28 03:10:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 28 03:09:30 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	69246	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	63814	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	30639	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	10366	5.099	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 102.000%		
7) Chloroethane-d5	1.672	69	8519	4.651	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 93.000%		
11) 1,1-Dichloroethene-d2	2.306	63	27023	3.566	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 71.400%		
20) 2-Butanone-d5	4.483	46	34298	46.146	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 92.300%		
24) Chloroform-d	5.068	84	49177	5.025	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 100.400%		
26) 1,2-Dichloroethane-d4	5.964	65	25143	4.701	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 94.000%		
32) Benzene-d6	5.982	84	77061	5.183	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 103.600%		
36) 1,2-Dichloropropane-d6	7.318	67	22213	4.683	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 93.600%		
41) Toluene-d8	8.653	98	76833	5.154	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 103.000%		
43) trans-1,3-Dichloroprop...	8.951	79	12206	5.606	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 112.200%		
46) 2-Hexanone-d5	9.390	63	35002	44.667	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 89.340%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	18331	5.056	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 101.200%		
66) 1,2-Dichlorobenzene-d4	12.323	152	28252	5.472	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 109.400%		
Target Compounds						Qvalue
16) Methylene chloride	2.800	84	2059	0.417	ug/L	# 63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102721\
Data File : VX024968.D
Acq On : 27 Oct 2021 12:24
Operator : JC/MD
Sample : M4304-14
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK001

Quant Time: Oct 28 03:10:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 28 03:09:30 2021
Response via : Initial Calibration

