

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX102722\  
 Data File : VX032426.D  
 Acq On : 27 Oct 2022 14:20  
 Operator : JC/MD  
 Sample : N5290-02  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 257248-8

Quant Time: Oct 28 07:12:57 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X102622W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 27 08:28:06 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 123718   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 198659   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 171923   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 77330    | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 80749    | 46.974   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 93.940%  |
| 35) Dibromofluoromethane    | 5.391  | 113   | 65162    | 45.984   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 91.960%  |
| 50) Toluene-d8              | 8.653  | 98    | 233857   | 47.460   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 94.920%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 85217    | 44.274   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 88.540%  |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX102722\  
Data File : VX032426.D  
Acq On : 27 Oct 2022 14:20  
Operator : JC/MD  
Sample : N5290-02  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
257248-8

Quant Time: Oct 28 07:12:57 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X102622W.M  
Quant Title : SW846 8260  
QLast Update : Thu Oct 27 08:28:06 2022  
Response via : Initial Calibration

