

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
 Data File : VX032430.D
 Acq On : 27 Oct 2022 15:53
 Operator : JC/MD
 Sample : N5264-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 28 07:13:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198808	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173197	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80458	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81062	48.069	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.140%
35) Dibromofluoromethane	5.391	113	64454	45.450	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.900%
50) Toluene-d8	8.653	98	233862	47.426	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.860%
62) 4-Bromofluorobenzene	11.079	95	87739	45.550	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.100%
Target Compounds						
20) Methylene Chloride	2.794	84	430	0.306	ug/l	Qvalue # 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
Data File : VX032430.D
Acq On : 27 Oct 2022 15:53
Operator : JC/MD
Sample : N5264-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 28 07:13:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

