

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
 Data File : VX032431.D
 Acq On : 27 Oct 2022 16:17
 Operator : JC/MD
 Sample : N5264-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 28 07:13:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	118399	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	193109	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172602	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77130	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79764	48.485	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.980%
35) Dibromofluoromethane	5.391	113	63296	45.951	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.900%
50) Toluene-d8	8.647	98	231255	48.281	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.560%
62) 4-Bromofluorobenzene	11.079	95	85268	45.574	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.140%
Target Compounds						
20) Methylene Chloride	2.788	84	454	0.331	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
Data File : VX032431.D
Acq On : 27 Oct 2022 16:17
Operator : JC/MD
Sample : N5264-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 28 07:13:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

