

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
 Data File : VX032433.D
 Acq On : 27 Oct 2022 17:03
 Operator : JC/MD
 Sample : N5314-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 221026025-03-TRIP-BLANK

Quant Time: Oct 28 07:13:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	123577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	201693	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83550	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82715	48.172	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.340%
35) Dibromofluoromethane	5.385	113	66304	46.086	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.180%
50) Toluene-d8	8.647	98	240254	48.025	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.060%
62) 4-Bromofluorobenzene	11.079	95	91485	46.816	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.640%
Target Compounds						
16) Acetone	2.386	43	16120	25.830	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
Data File : VX032433.D
Acq On : 27 Oct 2022 17:03
Operator : JC/MD
Sample : N5314-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
221026025-03-TRIP-BLANK

Quant Time: Oct 28 07:13:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

