

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
 Data File : VX032434.D
 Acq On : 27 Oct 2022 17:27
 Operator : JC/MD
 Sample : N5315-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 221026025-02-TRIP-BLANK

Quant Time: Oct 28 07:14:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	123822	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	201691	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177891	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81866	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82500	47.952	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.900%		
35) Dibromofluoromethane	5.385	113	65807	45.741	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.480%		
50) Toluene-d8	8.653	98	240320	48.039	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.080%		
62) 4-Bromofluorobenzene	11.079	95	88795	45.439	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.880%		
Target Compounds						
16) Acetone	2.386	43	19135	30.601	ug/l	99
30) Chloroform	5.099	83	1983	0.734	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
Data File : VX032434.D
Acq On : 27 Oct 2022 17:27
Operator : JC/MD
Sample : N5315-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
221026025-02-TRIP-BLANK

Quant Time: Oct 28 07:14:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

