

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
 Data File : VX032436.D
 Acq On : 27 Oct 2022 18:13
 Operator : JC/MD
 Sample : N5329-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BNM-1

Quant Time: Oct 28 07:14:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	119562	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198704	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175092	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81165	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81052	48.789	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.580%
35) Dibromofluoromethane	5.385	113	64045	45.186	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.380%
50) Toluene-d8	8.653	98	234293	47.538	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.080%
62) 4-Bromofluorobenzene	11.079	95	87118	45.251	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.500%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1418	2.348	ug/l #	85
64) Tetrachloroethene	9.275	164	3151	2.459	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
Data File : VX032436.D
Acq On : 27 Oct 2022 18:13
Operator : JC/MD
Sample : N5329-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BNM-1

Quant Time: Oct 28 07:14:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

