

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
 Data File : VX032443.D
 Acq On : 27 Oct 2022 20:56
 Operator : JC/MD
 Sample : N5249-12DL 2X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-GEP-1-20221020DL

Quant Time: Oct 28 07:15:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126281	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208896	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85242	48.581	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.160%
35) Dibromofluoromethane	5.391	113	67590	45.360	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.720%
50) Toluene-d8	8.653	98	248706	48.000	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.000%
62) 4-Bromofluorobenzene	11.079	95	93675	46.283	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.560%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	544	0.372	ug/l #	81
32) 1,1,1-Trichloroethane	5.379	97	6422	2.597	ug/l	95
44) Trichloroethene	7.129	130	111368	67.374	ug/l	100
64) Tetrachloroethene	9.275	164	519	0.381	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
Data File : VX032443.D
Acq On : 27 Oct 2022 20:56
Operator : JC/MD
Sample : N5249-12DL 2X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-GEP-1-20221020DL

Quant Time: Oct 28 07:15:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

