

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
 Data File : VX032444.D
 Acq On : 27 Oct 2022 21:20
 Operator : JC/MD
 Sample : N5249-01DL
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-FD-02-20221019

Quant Time: Oct 28 07:15:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 08:28:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	129193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	209500	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186750	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86036	47.928	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.860%
35) Dibromofluoromethane	5.385	113	67808	45.375	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.760%
50) Toluene-d8	8.647	98	250113	48.133	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.260%
62) 4-Bromofluorobenzene	11.079	95	91632	45.143	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.280%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.794	84	2116	1.415	ug/l	87
25) 2-Butanone	4.587	43	1915	1.895	ug/l	96
27) cis-1,2-Dichloroethene	4.501	96	799	0.490	ug/l	63
32) 1,1,1-Trichloroethane	5.373	97	1424	0.563	ug/l #	24
44) Trichloroethene	7.123	130	23576	14.222	ug/l	100
64) Tetrachloroethene	9.275	164	1516	1.109	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102722\
Data File : VX032444.D
Acq On : 27 Oct 2022 21:20
Operator : JC/MD
Sample : N5249-01DL
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-FD-02-20221019

Quant Time: Oct 28 07:15:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 08:28:06 2022
Response via : Initial Calibration

