

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038594.D
 Acq On : 27 Oct 2023 11:58
 Operator : JC/MD
 Sample : 05097-01 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22

Quant Time: Oct 30 01:28:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	216315	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	365407	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	369114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	203753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	142219	40.876	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.760%
35) Dibromofluoromethane	5.385	113	130169	49.438	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.880%
50) Toluene-d8	8.647	98	475221	47.233	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.460%
62) 4-Bromofluorobenzene	11.079	95	191196	50.126	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.260%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	77640	32.905	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	176478	60.471	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038594.D
Acq On : 27 Oct 2023 11:58
Operator : JC/MD
Sample : 05097-01 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Oct 30 01:28:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

