

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038598.D
 Acq On : 27 Oct 2023 13:30
 Operator : JC/MD
 Sample : 05063-01RE 2000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-5iRE

Quant Time: Oct 30 01:29:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	221084	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	366781	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	374894	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	206947	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	156241	43.937	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.880%
35) Dibromofluoromethane	5.385	113	140908	53.316	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.640%
50) Toluene-d8	8.647	98	513288	50.825	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.640%
62) 4-Bromofluorobenzene	11.079	95	208391	54.429	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.860%
Target Compounds						
27) cis-1,2-Dichloroethene	4.489	96	198551	66.567	ug/l	77
44) Trichloroethene	7.129	130	115432	43.296	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038598.D
Acq On : 27 Oct 2023 13:30
Operator : JC/MD
Sample : 05063-01RE 2000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-5iRE

Quant Time: Oct 30 01:29:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

