

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038599.D
 Acq On : 27 Oct 2023 13:53
 Operator : JC/MD
 Sample : 05063-03RE 2000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1RE

Quant Time: Oct 30 01:30:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	220319	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	369456	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	381244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	210145	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	154480	43.593	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.180%
35) Dibromofluoromethane	5.391	113	139824	52.523	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.040%
50) Toluene-d8	8.647	98	501990	49.346	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.700%
62) 4-Bromofluorobenzene	11.079	95	202877	52.605	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.220%
Target Compounds						
27) cis-1,2-Dichloroethene	4.489	96	201497	67.789	ug/l	Qvalue 78
44) Trichloroethene	7.129	130	118112	43.980	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038599.D
Acq On : 27 Oct 2023 13:53
Operator : JC/MD
Sample : 05063-03RE 2000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1RE

Quant Time: Oct 30 01:30:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

