

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038600.D
 Acq On : 27 Oct 2023 14:16
 Operator : JC/MD
 Sample : 05097-10RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01RE

Quant Time: Oct 30 01:30:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	215406	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	363093	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	377557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	208307	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139751	40.336	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	80.680%
35) Dibromofluoromethane	5.385	113	125725	48.054	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.100%
50) Toluene-d8	8.646	98	456508	45.662	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.320%#
62) 4-Bromofluorobenzene	11.079	95	185884	49.043	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.080%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	5399	3.734	ug/l	97
20) Methylene Chloride	2.788	84	4181	1.537	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038600.D
Acq On : 27 Oct 2023 14:16
Operator : JC/MD
Sample : 05097-10RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01RE

Quant Time: Oct 30 01:30:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

