

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038602.D
 Acq On : 27 Oct 2023 15:03
 Operator : JC/MD
 Sample : 05063-11RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-11SRE

Quant Time: Oct 30 01:31:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	213870	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	359916	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	368016	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	200874	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	134219	39.018	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	78.040%
35) Dibromofluoromethane	5.385	113	120670	46.529	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.060%
50) Toluene-d8	8.647	98	442050	44.606	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.220%#
62) 4-Bromofluorobenzene	11.079	95	179034	47.653	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.300%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	2468	1.058	ug/l	99
16) Acetone	2.380	43	2981	2.076	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	69753	24.175	ug/l	77
44) Trichloroethene	7.135	130	3710	1.418	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038602.D
Acq On : 27 Oct 2023 15:03
Operator : JC/MD
Sample : 05063-11RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-11SRE

Quant Time: Oct 30 01:31:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

