

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038603.D
 Acq On : 27 Oct 2023 15:26
 Operator : JC/MD
 Sample : 05063-13RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-11IRE

Quant Time: Oct 30 01:31:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	209470	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	354427	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	364597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	201925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138835	41.207	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.420%
35) Dibromofluoromethane	5.385	113	124992	48.942	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	8.647	98	455688	46.694	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.380%
62) 4-Bromofluorobenzene	11.079	95	181964	49.183	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.360%
Target Compounds						
27) cis-1,2-Dichloroethene	4.495	96	15734	5.568	ug/l	77
44) Trichloroethene	7.129	130	363238	140.992	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038603.D
Acq On : 27 Oct 2023 15:26
Operator : JC/MD
Sample : 05063-13RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-11IRE

Quant Time: Oct 30 01:31:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

