

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038604.D
 Acq On : 27 Oct 2023 15:49
 Operator : JC/MD
 Sample : 05063-15RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-11DRE

Quant Time: Oct 30 01:31:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	206927	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	348245	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	362661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	203213	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	132193	39.718	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.440%
35) Dibromofluoromethane	5.385	113	119040	47.439	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.880%
50) Toluene-d8	8.647	98	432762	45.132	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.260%#
62) 4-Bromofluorobenzene	11.079	95	178344	49.060	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.120%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	3923	1.550	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038604.D
Acq On : 27 Oct 2023 15:49
Operator : JC/MD
Sample : 05063-15RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-11DRE

Quant Time: Oct 30 01:31:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

