

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038605.D
 Acq On : 27 Oct 2023 16:12
 Operator : JC/MD
 Sample : 05097-03RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-03SRE

Quant Time: Nov 06 08:22:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	210374	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	354382	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	364726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	202210	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135052	39.912	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.820%
35) Dibromofluoromethane	5.385	113	120311	47.115	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.240%
50) Toluene-d8	8.647	98	440332	45.126	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.260%#
62) 4-Bromofluorobenzene	11.079	95	177576	48.003	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.000%
Target Compounds						
44) Trichloroethene	7.135	130	975	0.378	ug/l	Qvalue 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038605.D
Acq On : 27 Oct 2023 16:12
Operator : JC/MD
Sample : 05097-03RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-03SRE

Quant Time: Nov 06 08:22:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

