

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038606.D
 Acq On : 27 Oct 2023 16:36
 Operator : JC/MD
 Sample : 05102-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 67S-TB01

Quant Time: Oct 30 01:32:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	213018	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	356310	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	366649	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	201931	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	136162	39.741	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.480%
35) Dibromofluoromethane	5.385	113	121689	47.397	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.800%
50) Toluene-d8	8.646	98	443128	45.167	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.340%#
62) 4-Bromofluorobenzene	11.079	95	179431	48.242	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.480%
Target Compounds						
16) Acetone	2.379	43	17135	11.983	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038606.D
Acq On : 27 Oct 2023 16:36
Operator : JC/MD
Sample : 05102-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
67S-TB01

Quant Time: Oct 30 01:32:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

