

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038608.D
 Acq On : 27 Oct 2023 17:22
 Operator : JC/MD
 Sample : 05097-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-03i

Quant Time: Oct 30 01:33:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	217078	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	368134	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	376182	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	215671	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	147542	42.257	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.520%
35) Dibromofluoromethane	5.385	113	132935	50.114	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.220%
50) Toluene-d8	8.647	98	486081	47.954	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.900%
62) 4-Bromofluorobenzene	11.079	95	200204	52.098	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.200%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1452	0.530	ug/l	# 79
44) Trichloroethene	7.135	130	3688	1.378	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038608.D
Acq On : 27 Oct 2023 17:22
Operator : JC/MD
Sample : 05097-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-03i

Quant Time: Oct 30 01:33:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

