

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038609.D
 Acq On : 27 Oct 2023 17:45
 Operator : JC/MD
 Sample : 05108-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NWB-1965

Quant Time: Oct 30 01:33:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	208718	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	346707	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	359282	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	200906	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	151120	45.015	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.040%
35) Dibromofluoromethane	5.385	113	136994	54.836	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.680%
50) Toluene-d8	8.647	98	484147	50.715	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.440%
62) 4-Bromofluorobenzene	11.079	95	195563	54.036	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.080%
Target Compounds						
16) Acetone	2.380	43	15195	10.845	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038609.D
Acq On : 27 Oct 2023 17:45
Operator : JC/MD
Sample : 05108-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NWB-1965

Quant Time: Oct 30 01:33:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

