

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
 Data File : VX038611.D
 Acq On : 27 Oct 2023 18:32
 Operator : JC/MD
 Sample : 05107-17 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1072

Quant Time: Oct 30 01:34:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	234718	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	385516	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	397670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	225706	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	163702	43.361	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.720%
35) Dibromofluoromethane	5.385	113	147030	52.929	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.860%
50) Toluene-d8	8.647	98	531877	50.106	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.220%
62) 4-Bromofluorobenzene	11.079	95	223607	55.565	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102723\
Data File : VX038611.D
Acq On : 27 Oct 2023 18:32
Operator : JC/MD
Sample : 05107-17 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1072

Quant Time: Oct 30 01:34:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

