

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102725\
 Data File : VX048378.D
 Acq On : 27 Oct 2025 17:26
 Operator : JC/MD
 Sample : Q3447-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P001-TW3-251023-01

Quant Time: Oct 28 01:57:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	104175	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	207774	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	217355	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	111222	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	78485	53.734	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.460%	
35) Dibromofluoromethane	5.373	113	55337	39.012	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 78.020%	
50) Toluene-d8	8.635	98	230634	46.388	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 92.780%	
62) 4-Bromofluorobenzene	11.061	95	101648	53.741	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.480%	
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	16829	33.586	ug/l	98
18) Methyl Acetate	2.703	43	3172	3.031	ug/l	96
20) Methylene Chloride	2.794	84	6013	4.925	ug/l	96
43) Isopropyl Acetate	6.336	43	4091	1.490	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102725\
Data File : VX048378.D
Acq On : 27 Oct 2025 17:26
Operator : JC/MD
Sample : Q3447-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P001-TW3-251023-01

Quant Time: Oct 28 01:57:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

