

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102725\
 Data File : VX048379.D
 Acq On : 27 Oct 2025 17:47
 Operator : JC/MD
 Sample : Q3447-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P001-TW4-251023-01

Quant Time: Oct 28 01:58:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	121486	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	236810	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	246825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	128536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	92691	54.418	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.840%	
35) Dibromofluoromethane	5.379	113	62397	38.596	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 77.200%	
50) Toluene-d8	8.635	98	267256	47.163	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 94.320%	
62) 4-Bromofluorobenzene	11.061	95	118046	54.758	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.520%	
Target Compounds						Qvalue
16) Acetone	2.373	43	26248	44.919	ug/l	98
17) Carbon Disulfide	2.520	76	125382	31.192	ug/l	99
18) Methyl Acetate	2.703	43	2047	1.677	ug/l	91
20) Methylene Chloride	2.788	84	10430	7.325	ug/l	95
43) Isopropyl Acetate	6.324	43	10658	3.407	ug/l	96
52) Toluene	8.708	92	23620	5.573	ug/l	98
67) Ethyl Benzene	10.177	91	12668	1.340	ug/l	91
95) Naphthalene	13.756	128	8947	1.199	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102725\
Data File : VX048379.D
Acq On : 27 Oct 2025 17:47
Operator : JC/MD
Sample : Q3447-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P001-TW4-251023-01

Quant Time: Oct 28 01:58:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

