

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102819\
 Data File : VX013251.D
 Acq On : 28 Oct 2019 16:34
 Operator : JC/SP
 Sample : K5575-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Oct 29 15:02:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	69454	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	103873	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	101757	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	63070	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	80466	55.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.78%	
35) Dibromofluoromethane	5.49	113	62471	53.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.86%	
50) Toluene-d8	8.71	98	219064	51.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.06%	
62) 4-Bromofluorobenzene	11.14	95	81746	48.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.22%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102819\
Data File : VX013251.D
Acq On : 28 Oct 2019 16:34
Operator : JC/SP
Sample : K5575-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Oct 29 15:02:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

