

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102819\
 Data File : VX013254.D
 Acq On : 28 Oct 2019 17:44
 Operator : JC/SP
 Sample : K5575-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-3

Quant Time: Oct 29 15:10:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	69403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	102038	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	101440	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	64990	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	80953	55.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.54%	
35) Dibromofluoromethane	5.48	113	60612	53.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.54%	
50) Toluene-d8	8.71	98	214447	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
62) 4-Bromofluorobenzene	11.14	95	77288	46.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.62%	
Target Compounds						
16) Acetone	2.44	43	5593	8.178	ug/l	99
52) Toluene	8.78	92	1305	0.384	ug/l	88
68) m/p-Xylenes	10.35	106	870	0.362	ug/l	97
69) o-Xylene	10.69	106	810	0.351	ug/l	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102819\
Data File : VX013254.D
Acq On : 28 Oct 2019 17:44
Operator : JC/SP
Sample : K5575-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-3

Quant Time: Oct 29 15:10:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

