

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102819\
 Data File : VX013239.D
 Acq On : 28 Oct 2019 10:40
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 29 14:53:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	1.009	0.923	8.5	78	0.00
3 P	Chloromethane	0.886	0.742	16.3	75	0.00
4 C	Vinyl Chloride	0.908	0.783	13.8#	75	0.00
5 T	Bromomethane	0.481	0.425	11.6	81	0.00
6 T	Chloroethane	0.586	0.505	13.8	79	0.00
7 T	Trichlorofluoromethane	1.527	1.442	5.6	84	0.00
8 T	Diethyl Ether	0.491	0.482	1.8	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.887	0.855	3.6	80	0.00
10 T	Methyl Iodide	1.017	0.983	3.3	73	0.00
11 T	Tert butyl alcohol	0.241	0.230	4.6	85	0.00
12 CM	1,1-Dichloroethene	0.852	0.782	8.2#	79	0.00
13 T	Acrolein	0.123	0.107	13.0	75	0.00
14 T	Allyl chloride	1.341	1.219	9.1	79	0.00
15 T	Acrylonitrile	0.461	0.436	5.4	82	0.00
16 T	Acetone	0.493	0.571	-15.8	95	0.00
17 T	Carbon Disulfide	2.290	1.999	12.7	76	0.00
18 T	Methyl Acetate	1.129	1.040	7.9	81	0.00
19 T	Methyl tert-butyl Ether	2.999	2.979	0.7	86	0.00
20 T	Methylene Chloride	0.990	0.910	8.1	82	0.00
21 T	trans-1,2-Dichloroethene	0.945	0.855	9.5	80	0.00
22 T	Diisopropyl ether	2.622	2.509	4.3	83	0.00
23 T	Vinyl Acetate	2.229	2.218	0.5	83	0.00
24 P	1,1-Dichloroethane	1.613	1.513	6.2	82	0.00
25 T	2-Butanone	0.661	0.659	0.3	85	0.00
26 T	2,2-Dichloropropane	1.490	1.463	1.8	83	0.00
27 T	cis-1,2-Dichloroethene	1.043	0.989	5.2	83	0.00
28 T	Bromochloromethane	0.416	0.405	2.6	78	0.00
29 T	Tetrahydrofuran	0.398	0.361	9.3	80	0.00
30 C	Chloroform	1.717	1.689	1.6#	85	0.00
31 T	Cyclohexane	1.352	1.270	6.1	77	0.00
32 T	1,1,1-Trichloroethane	1.527	1.543	-1.0	84	0.00
33 S	1,2-Dichloroethane-d4	1.046	1.037	0.9	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.558	0.570	-2.2	83	0.00
36 T	1,1-Dichloropropene	0.896	0.854	4.7	81	0.00
37 T	Ethyl Acetate	0.833	0.809	2.9	83	0.00
38 T	Carbon Tetrachloride	0.924	0.960	-3.9	87	0.00
39 T	Methylcyclohexane	1.070	1.021	4.6	78	0.00
40 TM	Benzene	2.511	2.420	3.6	81	0.00
41 T	Methacrylonitrile	0.447	0.450	-0.7	83	-0.01
42 TM	1,2-Dichloroethane	0.947	1.010	-6.7	88	0.00
43 T	Isopropyl Acetate	1.383	1.379	0.3	82	0.00
44 TM	Trichloroethene	0.713	0.697	2.2	82	0.00
45 C	1,2-Dichloropropane	0.629	0.616	2.1#	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX102819\
 Data File : VX013239.D
 Acq On : 28 Oct 2019 10:40
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 29 14:53:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.453	0.460	-1.5	86	0.00
47 T	Bromodichloromethane	0.883	0.962	-8.9	90	0.00
48 T	Methyl methacrylate	0.684	0.680	0.6	84	0.00
49 T	1,4-Dioxane	0.017	0.017	0.0	84	0.00
50 S	Toluene-d8	2.066	2.053	0.6	81	0.00
51 T	4-Methyl-2-Pentanone	0.843	0.831	1.4	83	0.00
52 CM	Toluene	1.664	1.611	3.2#	83	0.00
53 T	t-1,3-Dichloropropene	0.939	1.054	-12.2	87	0.00
54 T	cis-1,3-Dichloropropene	1.027	1.097	-6.8	84	0.00
55 T	1,1,2-Trichloroethane	0.658	0.667	-1.4	86	0.00
56 T	Ethyl methacrylate	0.990	1.042	-5.3	86	0.00
57 T	1,3-Dichloropropane	1.110	1.108	0.2	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.386	0.419	-8.5	84	0.00
59 T	2-Hexanone	0.650	0.667	-2.6	83	0.00
60 T	Dibromochloromethane	0.690	0.782	-13.3	90	0.00
61 T	1,2-Dibromoethane	0.699	0.719	-2.9	86	0.00
62 S	4-Bromofluorobenzene	0.818	0.837	-2.3	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.683	0.668	2.2	84	0.00
65 PM	Chlorobenzene	1.750	1.704	2.6	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.647	0.668	-3.2	88	0.00
67 C	Ethyl Benzene	3.133	3.063	2.2#	83	0.00
68 T	m/p-Xylenes	1.184	1.147	3.1	83	0.00
69 T	o-Xylene	1.136	1.127	0.8	85	0.00
70 T	Styrene	1.956	1.964	-0.4	83	0.00
71 P	Bromoform	0.472	0.558	-18.2	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	4.540	4.425	2.5	85	0.00
74 T	N-amyl acetate	1.666	1.669	-0.2	83	0.00
75 P	1,1,2,2-Tetrachloroethane	1.379	1.355	1.7	84	0.00
76 T	1,2,3-Trichloropropane	1.205	1.363	-13.1	106	0.00
77 T	Bromobenzene	1.088	1.056	2.9	85	0.00
78 T	n-propylbenzene	4.932	4.929	0.1	85	0.00
79 T	2-Chlorotoluene	3.116	2.964	4.9	82	0.00
80 T	1,3,5-Trimethylbenzene	3.774	3.530	6.5	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.439	0.485	-10.5	91	0.00
82 T	4-Chlorotoluene	3.531	3.442	2.5	84	0.00
83 T	tert-Butylbenzene	3.734	3.500	6.3	82	0.00
84 T	1,2,4-Trimethylbenzene	3.807	3.690	3.1	82	0.00
85 T	sec-Butylbenzene	4.249	3.966	6.7	79	0.00
86 T	p-Isopropyltoluene	3.992	3.749	6.1	77	0.00
87 T	1,3-Dichlorobenzene	1.966	1.855	5.6	79	0.00
88 T	1,4-Dichlorobenzene	2.002	1.857	7.2	81	0.00
89 T	n-Butylbenzene	3.266	3.316	-1.5	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102819\
Data File : VX013239.D
Acq On : 28 Oct 2019 10:40
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 29 14:53:24 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.669	0.686	-2.5	83	0.00
91 T	1,2-Dichlorobenzene	1.933	1.890	2.2	84	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.343	0.362	-5.5	88	0.00
93 T	1,2,4-Trichlorobenzene	1.217	1.231	-1.2	82	0.00
94 T	Hexachlorobutadiene	0.630	0.608	3.5	79	0.00
95 T	Naphthalene	3.736	3.898	-4.3	84	0.00
96 T	1,2,3-Trichlorobenzene	1.203	1.256	-4.4	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6