


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102819\
 Data File : VX013238.D
 Acq On : 28 Oct 2019 09:29
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Title : SW846 8260
 Last Update : Thu Oct 24 03:56:38 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1673

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	4778	PASS
75	95	30	60	52.7	15059	PASS
95	95	100	100	100.0	28549	PASS
96	95	5	9	7.3	2095	PASS
173	174	0.00	2	0.5	105	PASS
174	95	50	100	71.6	20439	PASS
175	174	5	9	7.1	1458	PASS
176	174	95	101	95.7	19570	PASS
177	176	5	9	7.4	1444	PASS