

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102820\
 Data File : VX019160.D
 Acq On : 28 Oct 2020 13:09
 Operator : JC/MD
 Sample : L4470-03DL 4X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW179D2-GW-20201016DL

Quant Time: Nov 03 11:09:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 27 13:55:07 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	319305	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	552914	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	514062	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	255862	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	209752	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
35) Dibromofluoromethane	5.46	113	171507	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
50) Toluene-d8	8.70	98	673688	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
62) 4-Bromofluorobenzene	11.12	95	254649	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	23858	5.994	ug/l	90
44) Trichloroethene	7.19	130	169817	40.960	ug/l	95
64) Tetrachloroethene	9.32	164	12478	3.186	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102820\
Data File : VX019160.D
Acq On : 28 Oct 2020 13:09
Operator : JC/MD
Sample : L4470-03DL 4X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW179D2-GW-20201016DL

Quant Time: Nov 03 11:09:00 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102720W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 27 13:55:07 2020
Response via : Initial Calibration

